

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041019\
 Data File : VD061726.D
 Acq On : 10 Apr 2019 21:20
 Operator : VA/AP
 Sample : K2338-10RE
 Misc : 5.39g/5ml/MSVOA D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-2-DEEPR

Quant Time: Apr 11 01:09:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 08 10:04:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	6526	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	8.23	114	12027	50.00	ug/l	0.02
63) Chlorobenzene-d5	12.37	117	13411	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	6355	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	4393	88.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	177.62%	
35) Dibromofluoromethane	6.93	113	4483	53.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.60%	
50) Toluene-d8	10.41	98	11345	54.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.40%	
62) 4-Bromofluorobenzene	13.56	95	5044	76.89	ug/l	0.01
Spiked Amount 50.000			Recovery	=	153.78%	
Target Compounds						
					Qvalue	
5) Bromomethane	2.14	94	222	18.041	ug/l #	12
16) Acetone	3.21	43	5761	389.592	ug/l #	51
20) Methylene Chloride	3.82	84	34817	414.928	ug/l	96
81) trans-1,4-Dichloro-2-buten	13.56	75	1597	60.869	ug/l #	20

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041019\
Data File : VD061726.D
Acq On : 10 Apr 2019 21:20
Operator : VA/AP
Sample : K2338-10RE
Misc : 5.39g/5ml/MSVOA D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-2-DEEPR

Quant Time: Apr 11 01:09:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 08 10:04:03 2019
Response via : Initial Calibration

