

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041023\
 Data File : VD075616.D
 Acq On : 10 Apr 2023 19:51
 Operator : KP/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Apr 12 09:57:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041023S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 12 09:55:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.881	168	165716	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	406139	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	437106	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	204639	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	89477	65.818	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.640%
35) Dibromofluoromethane	7.810	113	138159	58.045	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.080%
50) Toluene-d8	10.269	98	484114	51.854	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.700%
62) 4-Bromofluorobenzene	12.575	95	173799	63.278	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	126.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041023\
 Data File : VD075616.D
 Acq On : 10 Apr 2023 19:51
 Operator : KP/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Apr 12 09:57:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041023S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 12 09:55:43 2023
 Response via : Initial Calibration

