

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041119\
 Data File : VD061756.D
 Acq On : 11 Apr 2019 22:15
 Operator : VA/AP
 Sample : K2203-05
 Misc : 6.66g/5ml/MSVOA D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-04-040919-A

Quant Time: Apr 12 03:49:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 08 10:04:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	457223	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.21	114	616941	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	141429	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	244874	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	201797	58.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.46%	
35) Dibromofluoromethane	6.92	113	325894	76.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	152.50%	
50) Toluene-d8	10.41	98	9516	0.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	1.78%	
62) 4-Bromofluorobenzene	13.56	95	256967	76.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	152.20%	
Target Compounds						
16) Acetone	3.21	43	14675	14.165	ug/l #	73
20) Methylene Chloride	3.81	84	67625	11.503	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041119\
Data File : VD061756.D
Acq On : 11 Apr 2019 22:15
Operator : VA/AP
Sample : K2203-05
Misc : 6.66g/5ml/MSVOA D/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-04-040919-A

Quant Time: Apr 12 03:49:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 08 10:04:03 2019
Response via : Initial Calibration

