

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041122\
 Data File : VD072581.D
 Acq On : 11 Apr 2022 12:05
 Operator : VA/SY
 Sample : N2317-07RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NORTH-2RE

Quant Time: Apr 12 01:02:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	153232	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	270949	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	273265	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	125201	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	125690	67.807	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.620%
35) Dibromofluoromethane	7.908	113	106935	50.785	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.560%
50) Toluene-d8	10.337	98	331713	45.320	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	90.640%
62) 4-Bromofluorobenzene	12.625	95	123468	43.193	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041122\
Data File : VD072581.D
Acq On : 11 Apr 2022 12:05
Operator : VA/SY
Sample : N2317-07RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NORTH-2RE

Quant Time: Apr 12 01:02:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

