

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041122\
 Data File : VD072583.D
 Acq On : 11 Apr 2022 13:01
 Operator : VA/SY
 Sample : N2317-21
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP2-A

Quant Time: Apr 12 01:02:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	228879	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	421018	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	423219	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	193152	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	181287	65.476	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.960%
35) Dibromofluoromethane	7.909	113	156988	47.981	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.960%
50) Toluene-d8	10.332	98	515829	45.354	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	90.700%
62) 4-Bromofluorobenzene	12.620	95	188610	42.463	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.920%
Target Compounds						
					Qvalue	
31) Cyclohexane	7.979	56	5288	1.461	ug/l #	31

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041122\
Data File : VD072583.D
Acq On : 11 Apr 2022 13:01
Operator : VA/SY
Sample : N2317-21
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP2-A

Quant Time: Apr 12 01:02:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

