

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041122\
 Data File : VD072587.D
 Acq On : 11 Apr 2022 14:54
 Operator : VA/SY
 Sample : N2335-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A350-CU

Quant Time: Apr 12 01:03:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	466771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	874827	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	842313	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	348314	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	288609	51.113	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.220%	
35) Dibromofluoromethane	7.908	113	285614	42.010	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 84.020%	
50) Toluene-d8	10.332	98	1105201	46.766	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 93.540%	
62) 4-Bromofluorobenzene	12.620	95	384688	41.680	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 83.360%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041122\
Data File : VD072587.D
Acq On : 11 Apr 2022 14:54
Operator : VA/SY
Sample : N2335-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
A350-CU

Quant Time: Apr 12 01:03:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

