

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041221\
 Data File : VD068834.D
 Acq On : 12 Apr 2021 16:09
 Operator : VA/SY
 Sample : M2008-01
 Misc : 3.86G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NANUET-ST-041221

Quant Time: Apr 13 01:10:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	278047	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	475591	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	505805	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	241863	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	116599	39.68	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	79.36%
35) Dibromofluoromethane	7.914	113	142632	46.11	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.22%
50) Toluene-d8	10.338	98	602377	51.62	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.24%
62) 4-Bromofluorobenzene	12.632	95	219148	56.54	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.08%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041221\
Data File : VD068834.D
Acq On : 12 Apr 2021 16:09
Operator : VA/SY
Sample : M2008-01
Misc : 3.86G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NANUET-ST-041221

Quant Time: Apr 13 01:10:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

