

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041320\
 Data File : VD065620.D
 Acq On : 13 Apr 2020 16:46
 Operator : VA/SY
 Sample : L2229-11
 Misc : 6.67G/5.00ml/MSVOA D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-40

Quant Time: Apr 14 03:58:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040920S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 10 01:39:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.99	168	465412	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	843296	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	818123	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	385615	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	262813	59.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.42%	
35) Dibromofluoromethane	7.92	113	292578	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
50) Toluene-d8	10.34	98	998744	46.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.84%	
62) 4-Bromofluorobenzene	12.64	95	335382	48.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.02%	
Target Compounds						
16) Acetone	4.17	43	56188	70.692	ug/l	97
17) Carbon Disulfide	4.40	76	18653	1.256	ug/l	96
25) 2-Butanone	7.23	43	11554	10.620	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041320\
Data File : VD065620.D
Acq On : 13 Apr 2020 16:46
Operator : VA/SY
Sample : L2229-11
Misc : 6.67G/5.00ml/MSVOA D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-40

Quant Time: Apr 14 03:58:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D040920S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 10 01:39:58 2020
Response via : Initial Calibration

