


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD041320\
 Data File : VD065613.D
 Acq On : 13 Apr 2020 10:11
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040920S.M
 Title : SW846 8260
 Last Update : Fri Apr 10 01:39:58 2020



AutoFind: Scans 1822, 1823, 1824; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	65954	PASS
75	95	30	60	47.8	183317	PASS
95	95	100	100	100.0	383637	PASS
96	95	5	9	6.5	24880	PASS
173	174	0.00	2	1.2	4230	PASS
174	95	50	100	94.0	360661	PASS
175	174	5	9	7.5	27064	PASS
176	174	95	101	97.3	351082	PASS
177	176	5	9	6.7	23648	PASS