


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD041519\
Data File : VD061794.D
Acq On : 15 Apr 2019 10:17
Operator : VA/AP
Sample : BFB
Misc : 5.00g/5ml/MSVOA D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Title : SW846 8260
Last Update : Tue Apr 16 08:53:01 2019



AutoFind: Scans 1234, 1235, 1236; Background Corrected with Scan 1229

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	14688	PASS
75	95	30	60	39.6	33370	PASS
95	95	100	100	100.0	84360	PASS
96	95	5	9	6.4	5429	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	87.0	73365	PASS
175	174	5	9	6.6	4877	PASS
176	174	95	101	99.1	72741	PASS
177	176	5	9	6.5	4709	PASS