

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041521\
 Data File : VD068851.D
 Acq On : 15 Apr 2021 12:53
 Operator : VA/SY
 Sample : M2029-01
 Misc : 5.74G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-01-041421

Quant Time: Apr 16 01:06:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	198516	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	340268	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	359709	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	184212	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	93964	44.79	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.58%
35) Dibromofluoromethane	7.914	113	115142	52.02	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.04%
50) Toluene-d8	10.338	98	429123	51.40	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.80%
62) 4-Bromofluorobenzene	12.626	95	178602	64.41	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	128.82%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041521\
Data File : VD068851.D
Acq On : 15 Apr 2021 12:53
Operator : VA/SY
Sample : M2029-01
Misc : 5.74G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-01-041421

Quant Time: Apr 16 01:06:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

