

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041521\
 Data File : VD068855.D
 Acq On : 15 Apr 2021 14:46
 Operator : VA/SY
 Sample : M2042-04
 Misc : 5.19G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FS-SB12A(41-42)

Quant Time: Apr 16 01:08:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	198317	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	362791	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	392813	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	199202	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	99412	47.43	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.86%
35) Dibromofluoromethane	7.914	113	115396	48.90	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.80%
50) Toluene-d8	10.338	98	456339	51.26	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.52%
62) 4-Bromofluorobenzene	12.626	95	182986	61.89	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	123.78%
Target Compounds						
16) Acetone	4.156	43	3083	10.62	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041521\
Data File : VD068855.D
Acq On : 15 Apr 2021 14:46
Operator : VA/SY
Sample : M2042-04
Misc : 5.19G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FS-SB12A(41-42)

Quant Time: Apr 16 01:08:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

