

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041521\
 Data File : VD068860.D
 Acq On : 15 Apr 2021 17:09
 Operator : VA/SY
 Sample : M2049-12
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-10-8.5-9.0

Quant Time: Apr 16 01:10:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	271658	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	454507	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	498855	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	273272	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	116378	40.54	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.08%
35) Dibromofluoromethane	7.914	113	150791	51.01	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.02%
50) Toluene-d8	10.338	98	576666	51.71	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.42%
62) 4-Bromofluorobenzene	12.626	95	239626	64.70	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	129.40%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041521\
Data File : VD068860.D
Acq On : 15 Apr 2021 17:09
Operator : VA/SY
Sample : M2049-12
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-10-8.5-9.0

Quant Time: Apr 16 01:10:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

