

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041619\
 Data File : VD061811.D
 Acq On : 16 Apr 2019 15:11
 Operator : VA/AP
 Sample : K2196-03
 Misc : 6.39g/5ml/MSVOA D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HD-02-041119

Quant Time: Apr 16 16:24:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 08:53:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	607259	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	8.23	114	997548	50.00	ug/l	-0.03
63) Chlorobenzene-d5	12.38	117	644666	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	14.52	152	202715	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	213732	42.00	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	84.00%	
35) Dibromofluoromethane	6.94	113	309654	41.09	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	82.18%	
50) Toluene-d8	10.42	98	521034	29.18	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	58.36%	
62) 4-Bromofluorobenzene	13.56	95	177365	24.51	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	49.02%	
Target Compounds						
16) Acetone	3.23	43	21644	17.255	ug/l #	87
20) Methylene Chloride	3.84	84	88695	10.612	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041619\
Data File : VD061811.D
Acq On : 16 Apr 2019 15:11
Operator : VA/AP
Sample : K2196-03
Misc : 6.39g/5ml/MSVOA D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HD-02-041119

Quant Time: Apr 16 16:24:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 08:53:01 2019
Response via : Initial Calibration

