

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041619\
 Data File : VD061817.D
 Acq On : 16 Apr 2019 18:27
 Operator : VA/AP
 Sample : K2377-06
 Misc : 5.69g/5ml/MSVOA D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-12-S2

Quant Time: Apr 17 05:43:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 08:53:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	415293	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	8.22	114	803227	50.00	ug/l	-0.04
63) Chlorobenzene-d5	12.37	117	490496	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	14.52	152	117283	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	211975	60.90	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	121.80%	
35) Dibromofluoromethane	6.92	113	313018	51.58	ug/l	-0.05
Spiked Amount 50.000			Recovery	=	103.16%	
50) Toluene-d8	10.41	98	612227	42.58	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	85.16%	
62) 4-Bromofluorobenzene	13.56	95	137481	23.59	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	47.18%	
Target Compounds						
16) Acetone	3.22	43	9686	11.291	ug/l	93
18) Methyl Acetate	3.65	43	8804	5.121	ug/l #	90
20) Methylene Chloride	3.81	84	56833	9.625	ug/l	95
52) Toluene	10.52	92	33532	3.321	ug/l	98
68) m/p-Xylenes	12.66	106	6356	1.157	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041619\
Data File : VD061817.D
Acq On : 16 Apr 2019 18:27
Operator : VA/AP
Sample : K2377-06
Misc : 5.69g/5ml/MSVOA D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-12-S2

Quant Time: Apr 17 05:43:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 08:53:01 2019
Response via : Initial Calibration

