

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041619\
 Data File : VD061818.D
 Acq On : 16 Apr 2019 18:54
 Operator : VA/AP
 Sample : K2376-02
 Misc : 2.59g/5ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-4-S2

Quant Time: Apr 17 01:12:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 08:53:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	123488	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	8.23	114	243660	50.00	ug/l	-0.03
63) Chlorobenzene-d5	12.37	117	158327	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	14.52	152	36140	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.47	65	123386	119.22	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	238.44%	
35) Dibromofluoromethane	6.93	113	161498	87.73	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	175.46%	
50) Toluene-d8	10.42	98	178218	40.86	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	81.72%	
62) 4-Bromofluorobenzene	13.56	95	45331	25.64	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	51.28%	
Target Compounds						
20) Methylene Chloride	3.82	84	29957	20.964	ug/l	96
40) Benzene	7.48	78	6385	1.336	ug/l #	87
52) Toluene	10.52	92	23946	7.817	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041619\
Data File : VD061818.D
Acq On : 16 Apr 2019 18:54
Operator : VA/AP
Sample : K2376-02
Misc : 2.59g/5ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-4-S2

Quant Time: Apr 17 01:12:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 08:53:01 2019
Response via : Initial Calibration

