

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041619\
 Data File : VD061825.D
 Acq On : 16 Apr 2019 23:33
 Operator : VA/AP
 Sample : VD0416SBL02
 Misc : 5.00µl/5ml/MSVOA D/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0416SBL02

Quant Time: Apr 17 02:06:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 08:53:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	614798	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	8.22	114	946693	50.00	ug/l	-0.04
63) Chlorobenzene-d5	12.37	117	821980	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	14.52	152	361926	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	280386	54.42	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	108.84%	
35) Dibromofluoromethane	6.93	113	409909	57.31	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	114.62%	
50) Toluene-d8	10.42	98	992809	58.58	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	117.16%	
62) 4-Bromofluorobenzene	13.56	95	372984	54.30	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	108.60%	
Target Compounds						
20) Methylene Chloride	3.82	84	40327	1.983	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041619\
Data File : VD061825.D
Acq On : 16 Apr 2019 23:33
Operator : VA/AP
Sample : VD0416SBL02
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0416SBL02

Quant Time: Apr 17 02:06:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 08:53:01 2019
Response via : Initial Calibration

