

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041619\
 Data File : VD061835.D
 Acq On : 17 Apr 2019 5:07
 Operator : VA/AP
 Sample : K2423-01
 Misc : 4.10g/5ml/MSVOA D/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP1-A

Quant Time: Apr 17 07:12:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 08:53:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	398314	50.00	ug/l	-0.05
34) 1,4-Difluorobenzene	8.22	114	701412	50.00	ug/l	-0.04
63) Chlorobenzene-d5	12.37	117	624554	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	14.51	152	287447	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	201107	60.24	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	120.48%	
35) Dibromofluoromethane	6.92	113	294084	55.50	ug/l	-0.05
Spiked Amount 50.000			Recovery	=	111.00%	
50) Toluene-d8	10.41	98	584990	46.59	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	93.18%	
62) 4-Bromofluorobenzene	13.55	95	263124	51.70	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	103.40%	
Target Compounds						
16) Acetone	3.22	43	51405	62.479	ug/l	99
20) Methylene Chloride	3.82	84	53183	9.268	ug/l	98
52) Toluene	10.51	92	33346	3.782	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041619\
Data File : VD061835.D
Acq On : 17 Apr 2019 5:07
Operator : VA/AP
Sample : K2423-01
Misc : 4.10g/5ml/MSVOA D/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP1-A

Quant Time: Apr 17 07:12:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 08:53:01 2019
Response via : Initial Calibration

