

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041619\
 Data File : VD061842.D
 Acq On : 17 Apr 2019 8:22
 Operator : VA/AP
 Sample : K2423-15
 Misc : 5.77g/5ml/MSVOA D/SOIL
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP4-B

Quant Time: Apr 17 12:41:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 08:53:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	397062	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	8.22	114	717338	50.00	ug/l	-0.04
63) Chlorobenzene-d5	12.37	117	625758	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	14.52	152	274934	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	172042	51.70	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	103.40%	
35) Dibromofluoromethane	6.93	113	254238	46.91	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	93.82%	
50) Toluene-d8	10.41	98	572102	44.55	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	89.10%	
62) 4-Bromofluorobenzene	13.56	95	227893	43.79	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	87.58%	
Target Compounds						
16) Acetone	3.23	43	37526	45.754	ug/l #	89
20) Methylene Chloride	3.83	84	65928	12.755	ug/l	94
52) Toluene	10.51	92	18778	2.082	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041619\
Data File : VD061842.D
Acq On : 17 Apr 2019 8:22
Operator : VA/AP
Sample : K2423-15
Misc : 5.77g/5ml/MSVOA D/SOIL
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP4-B

Quant Time: Apr 17 12:41:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 08:53:01 2019
Response via : Initial Calibration

