


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD041619\
Data File : VD061803.D
Acq On : 16 Apr 2019 9:30
Operator : VA/AP
Sample : BFB
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Title : SW846 8260
Last Update : Tue Apr 16 08:53:01 2019



AutoFind: Scans 1232, 1233, 1234; Background Corrected with Scan 1227

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	15615	PASS
75	95	30	60	37.2	36196	PASS
95	95	100	100	100.0	97354	PASS
96	95	5	9	6.2	6061	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	88.4	86098	PASS
175	174	5	9	6.4	5499	PASS
176	174	95	101	98.3	84613	PASS
177	176	5	9	6.5	5537	PASS