

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041621\
 Data File : VD068878.D
 Acq On : 16 Apr 2021 16:22
 Operator : VA/SY
 Sample : M2049-18
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-12-13.5-14.0

Quant Time: Apr 16 23:50:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	130121	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	232226	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	266148	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	135382	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	67428	49.03	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 98.06%			
35) Dibromofluoromethane	7.914	113	77462	51.28	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 102.56%			
50) Toluene-d8	10.338	98	298652	52.41	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 104.82%			
62) 4-Bromofluorobenzene	12.632	95	114044	60.26	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 120.52%			
Target Compounds					Qvalue	
16) Acetone	4.156	43	4840	25.42	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041621\
Data File : VD068878.D
Acq On : 16 Apr 2021 16:22
Operator : VA/SY
Sample : M2049-18
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-12-13.5-14.0

Quant Time: Apr 16 23:50:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

