

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041621\
 Data File : VD068882.D
 Acq On : 16 Apr 2021 18:15
 Operator : VA/SY
 Sample : M2046-01
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-4

Quant Time: Apr 16 23:51:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	277098	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	478524	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	514866	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	260546	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	112788	38.51	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	77.02%
35) Dibromofluoromethane	7.914	113	144259	46.35	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.70%
50) Toluene-d8	10.338	98	593390	50.54	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.08%
62) 4-Bromofluorobenzene	12.632	95	234193	60.05	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	120.10%
Target Compounds						
78) n-propylbenzene	12.820	91	23946	1.04	ug/l	99
84) 1,2,4-Trimethylbenzene	13.267	105	29747	1.85	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041621\
Data File : VD068882.D
Acq On : 16 Apr 2021 18:15
Operator : VA/SY
Sample : M2046-01
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-4

Quant Time: Apr 16 23:51:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

