

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041723\
 Data File : VD075700.D
 Acq On : 17 Apr 2023 14:41
 Operator : KP/SY
 Sample : 02258-01RE
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20230196-1-2RE

Quant Time: Apr 18 01:18:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	171040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	349673	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	473591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	254755	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	64842	39.940	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	79.880%
35) Dibromofluoromethane	7.810	113	107678	45.783	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.560%
50) Toluene-d8	10.275	98	525888	55.836	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.680%
62) 4-Bromofluorobenzene	12.575	95	200759	80.195	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	160.400%#
Target Compounds						
					Qvalue	
16) Acetone	4.022	43	3903	13.123	ug/l	92
17) Carbon Disulfide	4.269	76	17039	2.718	ug/l	95
20) Methylene Chloride	4.805	84	41548	14.813	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041723\
Data File : VD075700.D
Acq On : 17 Apr 2023 14:41
Operator : KP/SY
Sample : 02258-01RE
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20230196-1-2RE

Quant Time: Apr 18 01:18:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

