

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041723\
 Data File : VD075701.D
 Acq On : 17 Apr 2023 15:10
 Operator : KP/SY
 Sample : 02341-01RE
 Misc : 6.41G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PH1-BOT-1RE

Quant Time: Apr 18 01:18:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	172368	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	349423	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	472605	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	262133	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	63447	38.779	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	77.560%
35) Dibromofluoromethane	7.811	113	104920	44.643	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.280%
50) Toluene-d8	10.275	98	465322	49.441	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.880%
62) 4-Bromofluorobenzene	12.581	95	201588	80.584	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	161.160%#
Target Compounds						
20) Methylene Chloride	4.811	84	34602	11.359	ug/l	Qvalue # 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041723\
Data File : VD075701.D
Acq On : 17 Apr 2023 15:10
Operator : KP/SY
Sample : 02341-01RE
Misc : 6.41G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PH1-BOT-1RE

Quant Time: Apr 18 01:18:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

