

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041723\
 Data File : VD075712.D
 Acq On : 17 Apr 2023 20:33
 Operator : KP/SY
 Sample : 02270-14
 Misc : 2.44G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-24-5-7-20230411

Quant Time: Apr 18 01:19:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	188010	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	359720	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	319539	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	154589	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	96719	54.197	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.400%
35) Dibromofluoromethane	7.810	113	118849	49.122	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.240%
50) Toluene-d8	10.269	98	417549	43.095	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	86.200%
62) 4-Bromofluorobenzene	12.575	95	100242	38.924	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.840%
Target Compounds						
20) Methylene Chloride	4.805	84	73175	26.794	ug/l	Qvalue # 72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041723\
Data File : VD075712.D
Acq On : 17 Apr 2023 20:33
Operator : KP/SY
Sample : 02270-14
Misc : 2.44G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-24-5-7-20230411

Quant Time: Apr 18 01:19:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

