

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041819\
 Data File : VD061877.D
 Acq On : 18 Apr 2019 16:35
 Operator : VA/AP
 Sample : K2397-10RE
 Misc : 3.57g/5ml/MSVOA D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-3-S2RE

Quant Time: Apr 18 17:07:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 18 13:49:04 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	392728	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.22	114	664687	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	471892	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	134725	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	204316	62.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.14%	
35) Dibromofluoromethane	6.92	113	282346	56.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.44%	
50) Toluene-d8	10.41	98	638563	53.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.34%	
62) 4-Bromofluorobenzene	13.56	95	179824	37.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.58%	
Target Compounds						
16) Acetone	3.21	43	17545	21.628	ug/l	94
20) Methylene Chloride	3.82	84	24785	1.717	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041819\
Data File : VD061877.D
Acq On : 18 Apr 2019 16:35
Operator : VA/AP
Sample : K2397-10RE
Misc : 3.57a/5ml/MSVOA D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-3-S2RE

Quant Time: Apr 18 17:07:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 18 13:49:04 2019
Response via : Initial Calibration

