

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041819\
 Data File : VD061895.D
 Acq On : 19 Apr 2019 4:33
 Operator : VA/AP
 Sample : K2442-14
 Misc : 5.04g/5ml/MSVOA D/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-21-D2

Quant Time: Apr 19 05:56:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 18 13:49:04 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	408389	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.23	114	668173	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	621828	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	237822	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	210993	61.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.30%	
35) Dibromofluoromethane	6.94	113	291460	57.74	ug/l	0.02
Spiked Amount 50.000			Recovery	=	115.48%	
50) Toluene-d8	10.42	98	662228	55.37	ug/l	0.02
Spiked Amount 50.000			Recovery	=	110.74%	
62) 4-Bromofluorobenzene	13.56	95	259351	53.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.00%	
Target Compounds						
16) Acetone	3.23	43	28989	34.365	ug/l	99
20) Methylene Chloride	3.83	84	45361	6.860	ug/l	98
52) Toluene	10.51	92	66073	7.866	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041819\
Data File : VD061895.D
Acq On : 19 Apr 2019 4:33
Operator : VA/AP
Sample : K2442-14
Misc : 5.04g/5ml/MSVOA D/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-21-D2

Quant Time: Apr 19 05:56:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 18 13:49:04 2019
Response via : Initial Calibration

