

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041819\
 Data File : VD061894.D
 Acq On : 19 Apr 2019 4:05
 Operator : VA/AP
 Sample : K2442-10
 Misc : 5.27g/5ml/MSVOA D/SOIL
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-21-S2

Quant Time: Apr 19 05:53:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 18 13:49:04 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	372880	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.22	114	602393	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	439332	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	168880	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	191818	61.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.76%	
35) Dibromofluoromethane	6.93	113	281782	61.92	ug/l	0.01
Spiked Amount 50.000			Recovery	=	123.84%	
50) Toluene-d8	10.42	98	528595	49.02	ug/l	0.01
Spiked Amount 50.000			Recovery	=	98.04%	
62) 4-Bromofluorobenzene	13.56	95	213593	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
Target Compounds						
16) Acetone	3.21	43	27661	35.913	ug/l	95
20) Methylene Chloride	3.82	84	32859	4.399	ug/l	94
52) Toluene	10.52	92	44725	5.906	ug/l	98
68) m/p-Xylenes	12.66	106	5788	1.176	ug/l	84


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041819\
Data File : VD061894.D
Acq On : 19 Apr 2019 4:05
Operator : VA/AP
Sample : K2442-10
Misc : 5.27g/5ml/MSVOA D/SOIL
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-21-S2

Quant Time: Apr 19 05:53:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 18 13:49:04 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.05 min Scan# 572

Delta R.T. 0.00 min

Lab File: VD061894.D

Acq: 19 Apr 2019 4:05

Instrument :

MSVOA_D

ClientSampleId :


TP-21-S2

Tot Ion:168 Resp: 372880

Ion Ratio Lower Upper




168 100

99 55.8 46.5 69.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.95 (98.65 to 99.65): VDC



#16

Acetone

Concen: 35.913 ug/l

RT: 3.21 min Scan# 182

Delta R.T. 0.00 min

Lab File: VD061894.D


Acq: 19 Apr 2019 4:05

Tgt Ion: 43 Resp: 27661

Ion Ratio Lower Upper



43 100


58 22.2 19.9 29.9



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 57.95 (57.65 to 58.65): VDC









#20
Methvlene Chloride
Concen: 4.399 ug/l
RT: 3.82 min Scan# 244
Delta R.T. 0.01 min
Lab File: VD061894.D
Acq: 19 Apr 2019 4:05

Instrument :
MSVOA_D
ClientSampleId :
TP-21-S2

Tot Ion: 84 Resp: 32859
Ion Ratio Lower Upper
84 100
49 124.6 94.0 141.0
51 39.2 29.0 43.4
86 67.5 49.8 74.6




Abundance Ion 83.95 (83.65 to 84.65): VDC
Ion 48.90 (48.60 to 49.60): VDC
Ion 50.90 (50.60 to 51.60): VDC
Ion 85.90 (85.60 to 86.60): VDC






#33
1,2-Dichloroethane-d4
Concen: N.D. ug/l
RT: 7.46 min Scan# 613
Delta R.T. 0.00 min
Lab File: VD061894.D
Acq: 19 Apr 2019 4:05

Tgt Ion: 65 Resp: 191818
Ion Ratio Lower Upper
65 100
67 52.0 0.0 98.0



Abundance Ion 64.95 (64.65 to 65.65): VDC
Ion 66.90 (66.60 to 67.60): VDC





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.22 min Scan# 691

Delta R.T. 0.00 min

Lab File: VD061894.D

Acq: 19 Apr 2019 4:05

Instrument :

MSVOA_D

ClientSampleId :

TP-21-S2


Tot Ion:114 Resp: 602393

Ion Ratio Lower Upper

114 100

63 16.6 0.0 32.4




88 16.8 0.0 33.6



Abundance Ion 114.00 (113.70 to 114.70): V

Ion 62.95 (62.65 to 63.65): V

Ion 87.90 (87.60 to 88.60): V



#35

Dibromofluoromethane

Concen: 50.000 ug/l

RT: 6.93 min Scan# 560

Delta R.T. 0.01 min

Lab File: VD061894.D

Acq: 19 Apr 2019 4:05


Tgt Ion:113 Resp: 281782

Ion Ratio Lower Upper

113 100

111 107.5 78.5 117.7



192 15.5 13.0 19.4




Abundance Ion 112.90 (112.60 to 113.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 191.80 (191.50 to 192.50): V





#50

Toluene-d8

Concen: N.D. ug/l

RT: 10.42 min Scan# 914

Delta R.T. 0.01 min

Lab File: VD061894.D

Acq: 19 Apr 2019 4:05

Instrument :

MSVOA_D

ClientSampleId :


TP-21-S2

Tot Ion: 98 Resp: 528595




Ion Ratio Lower Upper

98 100

100 68.8 53.8 80.8



Abundance Ion 98.05 (97.75 to 98.75): VDC
Ion 100.05 (99.75 to 100.75): VDC



#52

Toluene

Concen: 5.906 ug/l

RT: 10.52 min Scan# 924

Delta R.T. 0.01 min

Lab File: VD061894.D


Acq: 19 Apr 2019 4:05

Tgt Ion: 92 Resp: 44725



Ion Ratio Lower Upper


92 100

91 174.0 136.6 204.8



Abundance Ion 92.00 (91.70 to 92.70): VDC
Ion 91.00 (90.70 to 91.70): VDC





#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 13.56 min Scan# 1233

Delta R.T. 0.00 min

Lab File: VD061894.D

Acq: 19 Apr 2019 4:05

Instrument :

MSVOA_D

ClientSampleId :

TP-21-S2


Tot Ion: 95 Resp: 213593

Ion Ratio Lower Upper

95 100

174 82.3 0.0 181.8




176 80.9 0.0 174.4



Abundance Ion 95.00 (94.70 to 95.70): V

Ion 173.90 (173.60 to 174.60): V

Ion 175.90 (175.60 to 176.60): V



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 12.37 min Scan# 1112

Delta R.T. 0.00 min

Lab File: VD061894.D

Acq: 19 Apr 2019 4:05


Tgt Ion: 117 Resp: 439332

Ion Ratio Lower Upper

117 100

82 47.7 36.2 54.2



119 31.4 26.2 39.4




Abundance Ion 116.95 (116.65 to 117.65): V

Ion 82.05 (81.75 to 82.75): V

Ion 118.95 (118.65 to 119.65): V









#68
m/p-Xvlenes
Concen: 1.176 ug/l
RT: 12.66 min Scan# 1142
Delta R.T. -0.01 min
Lab File: VD061894.D
Acq: 19 Apr 2019 4:05

Instrument :
MSVOA_D
ClientSampleId :
TP-21-S2

Tot Ion:106 Resp: 5788
Ion Ratio Lower Upper
106 100
91 180.7 164.0 246.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 14.52 min Scan# 1331
Delta R.T. 0.00 min
Lab File: VD061894.D
Acq: 19 Apr 2019 4:05

Tgt Ion:152 Resp: 168880
Ion Ratio Lower Upper
152 100
115 55.0 30.9 92.7
150 159.0 0.0 314.8

