

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
 Data File : VD072595.D
 Acq On : 18 Apr 2022 14:05
 Operator : VA/SY
 Sample : N2426-01
 Misc : 5.36G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PX-1

Quant Time: Apr 19 02:50:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	25487	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	45089	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	45628	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	19192	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	22973	74.511	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	149.020%
35) Dibromofluoromethane	7.908	113	18827	53.729	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.460%
50) Toluene-d8	10.332	98	52341	42.972	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	85.940%
62) 4-Bromofluorobenzene	12.620	95	19164	40.286	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
Data File : VD072595.D
Acq On : 18 Apr 2022 14:05
Operator : VA/SY
Sample : N2426-01
Misc : 5.36G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PX-1

Quant Time: Apr 19 02:50:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

