

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
 Data File : VD072598.D
 Acq On : 18 Apr 2022 15:29
 Operator : VA/SY
 Sample : N2426-04
 Misc : 5.22G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PX-4

Quant Time: Apr 19 02:50:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	188654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	360500	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	369959	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	170159	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	153291	67.170	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.340%
35) Dibromofluoromethane	7.908	113	135838	48.486	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.980%
50) Toluene-d8	10.331	98	430860	44.243	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.480%
62) 4-Bromofluorobenzene	12.620	95	160703	42.253	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
Data File : VD072598.D
Acq On : 18 Apr 2022 15:29
Operator : VA/SY
Sample : N2426-04
Misc : 5.22G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PX-4

Quant Time: Apr 19 02:50:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

