

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
 Data File : VD072600.D
 Acq On : 18 Apr 2022 16:26
 Operator : VA/SY
 Sample : N2444-01
 Misc : 4.92G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-1

Quant Time: Apr 19 02:51:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	416723	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	817333	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	767700	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	342224	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	240426	47.693	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.380%
35) Dibromofluoromethane	7.914	113	266528	41.961	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.920%
50) Toluene-d8	10.332	98	975985	44.204	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.400%
62) 4-Bromofluorobenzene	12.620	95	337974	39.195	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.380%
Target Compounds						
16) Acetone	4.156	43	13649	14.520	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
Data File : VD072600.D
Acq On : 18 Apr 2022 16:26
Operator : VA/SY
Sample : N2444-01
Misc : 4.92G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-1

Quant Time: Apr 19 02:51:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

