

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
 Data File : VD072602.D
 Acq On : 18 Apr 2022 17:22
 Operator : VA/SY
 Sample : N2444-03
 Misc : 4.66G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-3

Quant Time: Apr 19 02:51:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	374009	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	738020	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	725771	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	334484	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	246595	54.504	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.000%
35) Dibromofluoromethane	7.908	113	249964	43.582	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.160%
50) Toluene-d8	10.332	98	894142	44.849	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.700%
62) 4-Bromofluorobenzene	12.620	95	314362	40.374	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
Data File : VD072602.D
Acq On : 18 Apr 2022 17:22
Operator : VA/SY
Sample : N2444-03
Misc : 4.66G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-3

Quant Time: Apr 19 02:51:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

