

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
 Data File : VD072604.D
 Acq On : 18 Apr 2022 18:19
 Operator : VA/SY
 Sample : N2444-06
 Misc : 5.52G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-5

Quant Time: Apr 19 02:52:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	336061	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	675569	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	654834	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	296298	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	207707	51.093	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.180%
35) Dibromofluoromethane	7.909	113	218881	41.691	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.380%
50) Toluene-d8	10.332	98	800777	43.879	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	87.760%
62) 4-Bromofluorobenzene	12.620	95	284157	39.869	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	79.740%
Target Compounds					Qvalue	
16) Acetone	4.162	43	27687	36.524	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
Data File : VD072604.D
Acq On : 18 Apr 2022 18:19
Operator : VA/SY
Sample : N2444-06
Misc : 5.52G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-5

Quant Time: Apr 19 02:52:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

