

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
 Data File : VD072605.D
 Acq On : 18 Apr 2022 18:47
 Operator : VA/SY
 Sample : N2444-07
 Misc : 5.40G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-6

Quant Time: Apr 19 02:52:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	408279	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	827870	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	799802	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	358451	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	260652	52.775	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.560%
35) Dibromofluoromethane	7.908	113	269890	41.949	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.900%
50) Toluene-d8	10.332	98	1000265	44.727	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.460%
62) 4-Bromofluorobenzene	12.620	95	350208	40.097	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.200%
Target Compounds						
					Qvalue	
16) Acetone	4.155	43	34372	37.322	ug/l #	89
25) 2-Butanone	7.208	43	11023	8.654	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
Data File : VD072605.D
Acq On : 18 Apr 2022 18:47
Operator : VA/SY
Sample : N2444-07
Misc : 5.40G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-6

Quant Time: Apr 19 02:52:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

