

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
 Data File : VD072606.D
 Acq On : 18 Apr 2022 19:16
 Operator : VA/SY
 Sample : N2444-08
 Misc : 4.94G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-7

Quant Time: Apr 19 02:52:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	418388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	812955	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	784164	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	362465	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	257690	50.915	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.820%
35) Dibromofluoromethane	7.909	113	269981	42.733	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.460%
50) Toluene-d8	10.332	98	999051	45.492	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	90.980%
62) 4-Bromofluorobenzene	12.620	95	352661	41.118	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.240%
Target Compounds						
						Qvalue
16) Acetone	4.150	43	19250	20.397	ug/l	# 78
95) Naphthalene	15.385	128	499232	34.567	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
Data File : VD072606.D
Acq On : 18 Apr 2022 19:16
Operator : VA/SY
Sample : N2444-08
Misc : 4.94G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-7

Quant Time: Apr 19 02:52:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

