

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
 Data File : VD072608.D
 Acq On : 18 Apr 2022 20:12
 Operator : VA/SY
 Sample : N2444-10
 Misc : 5.26G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-10

Quant Time: Apr 19 02:53:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	384224	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	784987	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	764028	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	341619	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	246196	52.969	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.940%
35) Dibromofluoromethane	7.909	113	255809	41.933	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.860%
50) Toluene-d8	10.332	98	950261	44.812	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.620%
62) 4-Bromofluorobenzene	12.620	95	335615	40.525	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
Data File : VD072608.D
Acq On : 18 Apr 2022 20:12
Operator : VA/SY
Sample : N2444-10
Misc : 5.26G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-10

Quant Time: Apr 19 02:53:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

