

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
 Data File : VD072611.D
 Acq On : 18 Apr 2022 21:37
 Operator : VA/SY
 Sample : N2444-13
 Misc : 5.35G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-13

Quant Time: Apr 19 02:53:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	375444	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	771310	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	758835	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	344216	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	247485	54.491	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.980%	
35) Dibromofluoromethane	7.908	113	259866	43.353	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 86.700%	
50) Toluene-d8	10.332	98	948160	45.506	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 91.020%	
62) 4-Bromofluorobenzene	12.620	95	330524	40.618	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 81.240%	
Target Compounds						Qvalue
20) Methylene Chloride	4.967	84	32839	6.513	ug/l	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041822\
Data File : VD072611.D
Acq On : 18 Apr 2022 21:37
Operator : VA/SY
Sample : N2444-13
Misc : 5.35G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-13

Quant Time: Apr 19 02:53:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

