

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041823\
 Data File : VD075722.D
 Acq On : 18 Apr 2023 12:55
 Operator : KP/SY
 Sample : 02388-01
 Misc : 4.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CC-1

Quant Time: Apr 19 01:23:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	7420	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	15260	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	16536	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	8999	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.240	65	8358	118.671	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	237.340%	#
35) Dibromofluoromethane	7.805	113	222	2.163	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	4.320%	#
50) Toluene-d8	10.275	98	18513	45.041	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	90.080%	
62) 4-Bromofluorobenzene	12.575	95	6099	55.826	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	111.660%	
Target Compounds						
						Qvalue
16) Acetone	4.028	43	1742	135.010	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041823\
Data File : VD075722.D
Acq On : 18 Apr 2023 12:55
Operator : KP/SY
Sample : 02388-01
Misc : 4.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CC-1

Quant Time: Apr 19 01:23:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

