

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041823\
 Data File : VD075723.D
 Acq On : 18 Apr 2023 13:24
 Operator : KP/SY
 Sample : 02369-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-04-041423

Quant Time: Apr 19 01:23:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	14742	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	28120	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.586	117	26411	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	9244	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.233	65	13202	94.347	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	188.700%#
35) Dibromofluoromethane	7.804	113	13326	70.458	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	140.920%
50) Toluene-d8	10.269	98	33783	44.603	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.200%
62) 4-Bromofluorobenzene	12.574	95	8443	41.939	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041823\
Data File : VD075723.D
Acq On : 18 Apr 2023 13:24
Operator : KP/SY
Sample : 02369-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-04-041423

Quant Time: Apr 19 01:23:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

