

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041823\
 Data File : VD075737.D
 Acq On : 18 Apr 2023 20:10
 Operator : KP/SY
 Sample : 02388-01RE
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CC-1RE

Quant Time: Apr 19 01:25:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	165123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	342672	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	312740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	125434	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	108185	69.025	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	138.040%
35) Dibromofluoromethane	7.816	113	586	0.254	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.500%#
50) Toluene-d8	10.269	98	438362	47.494	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.980%
62) 4-Bromofluorobenzene	12.575	95	112189	45.731	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.460%
Target Compounds						
16) Acetone	4.022	43	9535	33.207	ug/l	Qvalue # 63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041823\
Data File : VD075737.D
Acq On : 18 Apr 2023 20:10
Operator : KP/SY
Sample : 02388-01RE
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CC-1RE

Quant Time: Apr 19 01:25:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

