

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041823\
 Data File : VD075716.D
 Acq On : 18 Apr 2023 09:23
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 19 01:20:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.465	0.491	-5.6	102	0.00
3 P	Chloromethane	0.839	0.738	12.0	85	0.00
4 C	Vinyl Chloride	1.113	1.066	4.2#	92	0.00
5 T	Bromomethane	0.869	0.793	8.7	90	0.00
6 T	Chloroethane	0.749	0.677	9.6	87	0.00
7 T	Trichlorofluoromethane	0.863	0.905	-4.9	101	0.00
8 T	Diethyl Ether	0.268	0.249	7.1	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.541	0.587	-8.5	104	0.00
10 T	Methyl Iodide	0.740	0.748	-1.1	91	0.00
11 T	Tert butyl alcohol	0.024	0.020	16.7	81	0.00
12 CM	1,1-Dichloroethene	0.560	0.583	-4.1#	98	0.00
13 T	Acrolein	0.048	0.050	-4.2	100	0.00
14 T	Allyl chloride	0.510	0.503	1.4	91	0.00
15 T	Acrylonitrile	0.100	0.098	2.0	92	0.00
16 T	Acetone	0.087	0.095	-9.2	94	0.00
17 T	Carbon Disulfide	1.833	1.749	4.6	92	0.00
18 T	Methyl Acetate	0.293	0.278	5.1	90	0.00
19 T	Methyl tert-butyl Ether	1.087	1.052	3.2	88	0.00
20 T	Methylene Chloride	0.800	0.650	18.8	90	0.00
21 T	trans-1,2-Dichloroethene	0.667	0.673	-0.9	95	0.00
22 T	Diisopropyl ether	1.153	1.168	-1.3	92	0.00
23 T	Vinyl Acetate	0.535	0.564	-5.4	92	0.00
24 P	1,1-Dichloroethane	0.955	0.947	0.8	94	0.00
25 T	2-Butanone	0.112	0.115	-2.7	94	0.00
26 T	2,2-Dichloropropane	0.866	0.840	3.0	94	0.00
27 T	cis-1,2-Dichloroethene	0.739	0.728	1.5	92	0.00
28 T	Bromochloromethane	0.302	0.292	3.3	94	0.00
29 T	Tetrahydrofuran	0.061	0.062	-1.6	92	0.00
30 C	Chloroform	1.083	1.079	0.4#	96	0.00
31 T	Cyclohexane	0.745	0.777	-4.3	101	0.00
32 T	1,1,1-Trichloroethane	0.883	0.909	-2.9	96	0.00
33 S	1,2-Dichloroethane-d4	0.475	0.427	10.1	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.336	0.332	1.2	91	0.00
36 T	1,1-Dichloropropene	0.464	0.495	-6.7	95	0.00
37 T	Ethyl Acetate	0.141	0.149	-5.7	98	0.00
38 T	Carbon Tetrachloride	0.409	0.460	-12.5	100	0.00
39 T	Methylcyclohexane	0.536	0.628	-17.2	100	0.00
40 TM	Benzene	1.442	1.514	-5.0	95	0.00
41 T	Methacrylonitrile	0.068	0.084	-23.5	121	0.00
42 TM	1,2-Dichloroethane	0.328	0.339	-3.4	92	0.00
43 T	Isopropyl Acetate	0.257	0.261	-1.6	90	0.00
44 TM	Trichloroethene	0.408	0.429	-5.1	95	0.00
45 C	1,2-Dichloropropane	0.323	0.337	-4.3#	95	0.00
46 T	Dibromomethane	0.205	0.208	-1.5	92	0.00
47 T	Bromodichloromethane	0.469	0.490	-4.5	95	0.00
48 T	Methyl methacrylate	0.115	0.123	-7.0	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041823\
 Data File : VD075716.D
 Acq On : 18 Apr 2023 09:23
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 19 01:20:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	87	0.00
50 S	Toluene-d8	1.347	1.334	1.0	81	0.00
51 T	4-Methyl-2-Pentanone	0.137	0.139	-1.5	81	0.00
52 CM	Toluene	0.959	0.991	-3.3#	83	0.00
53 T	t-1,3-Dichloropropene	0.450	0.443	1.6	82	0.00
54 T	cis-1,3-Dichloropropene	0.534	0.543	-1.7	83	0.00
55 T	1,1,2-Trichloroethane	0.297	0.294	1.0	85	0.00
56 T	Ethyl methacrylate	0.292	0.294	-0.7	79	0.00
57 T	1,3-Dichloropropane	0.451	0.447	0.9	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.121	3.2	78	0.00
59 T	2-Hexanone	0.097	0.107	-10.3	88	0.00
60 T	Dibromochloromethane	0.347	0.348	-0.3	85	0.00
61 T	1,2-Dibromoethane	0.283	0.270	4.6	81	0.00
62 S	4-Bromofluorobenzene	0.358	0.365	-2.0	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.360	0.382	-6.1	82	0.00
65 PM	Chlorobenzene	1.071	1.133	-5.8	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.368	0.398	-8.2	95	0.00
67 C	Ethyl Benzene	1.738	1.961	-12.8#	93	0.00
68 T	m/p-Xylenes	0.712	0.823	-15.6	96	0.00
69 T	o-Xylene	0.644	0.748	-16.1	94	0.00
70 T	Styrene	1.097	1.290	-17.6	93	0.00
71 P	Bromoform	0.192	0.212	-10.4	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.281	3.786	-15.4	95	0.00
74 T	N-amyl acetate	0.482	0.521	-8.1	92	0.00
75 P	1,1,2,2-Tetrachloroethane	0.637	0.659	-3.5	92	0.00
76 T	1,2,3-Trichloropropane	0.408	0.385	5.6	95	0.00
77 T	Bromobenzene	0.846	0.907	-7.2	95	0.00
78 T	n-propylbenzene	4.022	4.610	-14.6	95	0.00
79 T	2-Chlorotoluene	2.281	2.440	-7.0	94	0.00
80 T	1,3,5-Trimethylbenzene	2.736	3.097	-13.2	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.188	0.207	-10.1	91	0.00
82 T	4-Chlorotoluene	2.372	2.534	-6.8	93	0.00
83 T	tert-Butylbenzene	2.307	2.646	-14.7	96	0.00
84 T	1,2,4-Trimethylbenzene	2.745	3.083	-12.3	93	0.00
85 T	sec-Butylbenzene	3.558	4.163	-17.0	97	0.00
86 T	p-Isopropyltoluene	2.962	3.419	-15.4	95	0.00
87 T	1,3-Dichlorobenzene	1.760	1.879	-6.8	93	0.00
88 T	1,4-Dichlorobenzene	1.739	1.838	-5.7	94	0.00
89 T	n-Butylbenzene	2.729	3.146	-15.3	97	0.00
90 T	Hexachloroethane	0.589	0.645	-9.5	99	0.00
91 T	1,2-Dichlorobenzene	1.508	1.590	-5.4	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.085	0.081	4.7	87	0.00
93 T	1,2,4-Trichlorobenzene	0.877	0.880	-0.3	92	0.00
94 T	Hexachlorobutadiene	0.419	0.438	-4.5	99	0.00
95 T	Naphthalene	1.662	1.691	-1.7	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041823\
Data File : VD075716.D
Acq On : 18 Apr 2023 09:23
Operator : KP/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Apr 19 01:20:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.772	0.762	1.3	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6