



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041823\
 Data File : VD075715.D
 Acq On : 18 Apr 2023 08:53
 Operator : KP/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Title : SW846 8260
 Last Update : Thu Apr 13 14:16:29 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.6	9817	PASS
75	95	30	60	50.3	29720	PASS
95	95	100	100	100.0	59069	PASS
96	95	5	9	7.9	4646	PASS
173	174	0.00	2	0.7	394	PASS
174	95	50	100	93.7	55325	PASS
175	174	5	9	6.9	3811	PASS
176	174	95	101	95.0	52581	PASS
177	176	5	9	6.1	3205	PASS