

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041919\
 Data File : VD061908.D
 Acq On : 19 Apr 2019 14:16
 Operator : VA/AP
 Sample : K2422-03
 Misc : 5.95g/5ml/MSVOA D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP1-S

Quant Time: Apr 19 16:19:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 18 13:49:04 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	411146	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.23	114	661289	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	500386	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.53	152	179163	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	207816	60.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.62%	
35) Dibromofluoromethane	6.94	113	309136	61.88	ug/l	0.01
Spiked Amount 50.000			Recovery	=	123.76%	
50) Toluene-d8	10.42	98	574531	48.53	ug/l	0.01
Spiked Amount 50.000			Recovery	=	97.06%	
62) 4-Bromofluorobenzene	13.56	95	195653	40.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.56%	
Target Compounds						
16) Acetone	3.22	43	66360	78.139	ug/l	99
20) Methylene Chloride	3.83	84	64429	11.754	ug/l	93
52) Toluene	10.52	92	16104	1.937	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041919\
Data File : VD061908.D
Acq On : 19 Apr 2019 14:16
Operator : VA/AP
Sample : K2422-03
Misc : 5.95g/5ml/MSVOA D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP1-S

Quant Time: Apr 19 16:19:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 18 13:49:04 2019
Response via : Initial Calibration

