

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041921\
 Data File : VD068894.D
 Acq On : 19 Apr 2021 12:47
 Operator : VA/SY
 Sample : M2063-05
 Misc : 5.88G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP2-0-2

Quant Time: Apr 20 00:51:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	214049	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	367804	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	403383	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	198256	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	96193	42.52	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.04%
35) Dibromofluoromethane	7.920	113	114822	48.00	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.00%
50) Toluene-d8	10.344	98	463022	51.31	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.62%
62) 4-Bromofluorobenzene	12.632	95	188534	62.90	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	125.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041921\
Data File : VD068894.D
Acq On : 19 Apr 2021 12:47
Operator : VA/SY
Sample : M2063-05
Misc : 5.88G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP2-0-2

Quant Time: Apr 20 00:51:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

