

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041921\
 Data File : VD068903.D
 Acq On : 19 Apr 2021 17:02
 Operator : VA/SY
 Sample : M2077-05
 Misc : 5.16G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-20-14.5-15.0-DUP

Quant Time: Apr 20 00:53:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	18848	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.867	114	31690	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	43073	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	28183	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	13070	65.61	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 131.22%		
35) Dibromofluoromethane	7.920	113	11397	55.29	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 110.58%		
50) Toluene-d8	10.344	98	39499	50.80	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 101.60%		
62) 4-Bromofluorobenzene	12.626	95	19876	76.96	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 153.92%#		
Target Compounds						
16) Acetone	4.167	43	5554	201.35	ug/l	Qvalue # 83
73) Isopropylbenzene	12.479	105	26044	12.66	ug/l	98
84) 1,2,4-Trimethylbenzene	13.261	105	16026	9.22	ug/l	96
95) Naphthalene	15.396	128	916482	1002.57	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041921\
Data File : VD068903.D
Acq On : 19 Apr 2021 17:02
Operator : VA/SY
Sample : M2077-05
Misc : 5.16G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-20-14.5-15.0-DUP

Quant Time: Apr 20 00:53:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

