

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041921\
 Data File : VD068908.D
 Acq On : 19 Apr 2021 19:24
 Operator : VA/SY
 Sample : M2076-15
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-17-14.5-15.0

Quant Time: Apr 20 00:55:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	177058	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	311001	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	351063	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	175180	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	76791	41.04	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	82.08%
35) Dibromofluoromethane	7.914	113	98600	48.74	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.48%
50) Toluene-d8	10.338	98	394825	51.74	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.48%
62) 4-Bromofluorobenzene	12.632	95	156359	61.69	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	123.38%
Target Compounds						
16) Acetone	4.156	43	38742	149.51	ug/l	90
17) Carbon Disulfide	4.409	76	11532	2.01	ug/l	96
25) 2-Butanone	7.215	43	9838	26.83	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041921\
Data File : VD068908.D
Acq On : 19 Apr 2021 19:24
Operator : VA/SY
Sample : M2076-15
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-17-14.5-15.0

Quant Time: Apr 20 00:55:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

