

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041922\
 Data File : VD072624.D
 Acq On : 19 Apr 2022 14:54
 Operator : VA/SY
 Sample : N2426-01
 Misc : 5.37G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PX-1

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 04/20/2022
 Supervised By :Mahesh Dadoda 04/20/2022

Quant Time: Apr 20 05:03:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	196296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	384959	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	407480	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	195883	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	160104	67.424	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.840%
35) Dibromofluoromethane	7.914	113	141361	47.251	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.500%
50) Toluene-d8	10.332	98	472826	45.467	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	90.940%
62) 4-Bromofluorobenzene	12.620	95	184505	45.429	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.860%
Target Compounds						
16) Acetone	4.162	43	8129m	18.359	ug/l	
31) Cyclohexane	7.979	56	4889	1.575	ug/l #	55

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041922\
Data File : VD072624.D
Acq On : 19 Apr 2022 14:54
Operator : VA/SY
Sample : N2426-01
Misc : 5.37G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PX-1

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/20/2022
Supervised By :Mahesh Dadoda 04/20/2022

Quant Time: Apr 20 05:03:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

