

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041922\
 Data File : VD072627.D
 Acq On : 19 Apr 2022 16:18
 Operator : VA/SY
 Sample : N2442-05
 Misc : 4.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 124042

Quant Time: Apr 20 05:03:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	68282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	132982	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	141985	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	64251	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	68189	82.553	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	165.100%#
35) Dibromofluoromethane	7.909	113	56018	54.204	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.400%
50) Toluene-d8	10.332	98	159554	44.415	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.820%
62) 4-Bromofluorobenzene	12.620	95	59701	42.553	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	85.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041922\
Data File : VD072627.D
Acq On : 19 Apr 2022 16:18
Operator : VA/SY
Sample : N2442-05
Misc : 4.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
124042

Quant Time: Apr 20 05:03:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

