

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041922\
 Data File : VD072628.D
 Acq On : 19 Apr 2022 16:47
 Operator : VA/SY
 Sample : N2445-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 282

Quant Time: Apr 20 05:04:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	85729	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	169635	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	202036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	99534	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	114200	110.119	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 220.240%#	
35) Dibromofluoromethane	7.914	113	77575	58.845	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 117.680%	
50) Toluene-d8	10.332	98	213587	46.609	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 93.220%	
62) 4-Bromofluorobenzene	12.620	95	89839	50.199	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.400%	
Target Compounds						
16) Acetone	4.156	43	6209	32.108	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041922\
Data File : VD072628.D
Acq On : 19 Apr 2022 16:47
Operator : VA/SY
Sample : N2445-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
282

Quant Time: Apr 20 05:04:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

