

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041922\
 Data File : VD072630.D
 Acq On : 19 Apr 2022 17:44
 Operator : VA/SY
 Sample : N2449-01
 Misc : 5.34G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 22L091-A

Quant Time: Apr 20 05:04:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	435984	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	832207	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	795190	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	357503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	263947	50.046	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	100.100%	
35) Dibromofluoromethane	7.908	113	270991	41.901	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	83.800%	
50) Toluene-d8	10.332	98	991661	44.111	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	88.220%	
62) 4-Bromofluorobenzene	12.620	95	354062	40.327	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	80.660%	
Target Compounds						
16) Acetone	4.162	43	15955	16.224	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041922\
Data File : VD072630.D
Acq On : 19 Apr 2022 17:44
Operator : VA/SY
Sample : N2449-01
Misc : 5.34G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
22L091-A

Quant Time: Apr 20 05:04:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

